



Uniphos Envirotronic Pvt.Ltd.

C.I.N. U29297GJ2007PTC052025

A UPL Group Company

Nahuli, Via-Vapi, Tal. - Umbergaon,
Dist. Valsad. GUJARAT. INDIA - 396 108.
Tel.: +91-260-2730156
+91-9909994042, 7574839945
E-mail : uepl@uniphos-envirotronic.com
Website : www.uniphos-she.com



Safety & Environmental Monitoring Solutions

ISO 9001:2015 & ISO 14001:2015 CERTIFIED COMPANY BY BUREAU VERITAS

CALIBRATION REPORT

Calibration Report No : UEPL/QC/CAL/OR/202101225
Instrument : UNIPHOS KWIKALERT MARK - III
350(M) PHOSPHINE
Instrument Specification : Sr. No. : B200272 Type : Diffusion
Range : 0 - 20 ppm
Supplied to : For Safety S.A.S. - Argentina
Date of receipt of the Instrument : 15.01.2021
Calibration at : UNIPHOS ENVIROTRONIC PVT. LTD. Nahuli.
Calibration Gas used : Phosphine
Calibration Standard Applied : 10.80 ppm
Calibration Date : 15.01.2021
Next Calibration before : 15.01.2022

Calibration Check :

Sr. No.	Std. Gas Conc. Applied. (ppm)	Monitor Reading (ppm)
1	5.40	5.44
2	10.80	10.91

The results confirm that the performance of the instrument is within acceptable limits.

Calibrated by

Vinay Patel
Officer Tech

Checked by

Tejas Patel
Asst. Manager

FOR UNIPHOS ENVIROTRONIC PVT. LTD.

Ghanshyam Dubey
Dy. Manager R & D

Note :-

It is recommended to re-calibrate the instrument within a period of one year from the date of calibration. Please follow strictly all the instructions mentioned in the User Manual, to ensure accurate readings. During calibration Ambient Temperature 27.1 °C and Humidity 58 %RH

Regd. Office : 3-11, GIDC, VAPI 396 195, GUJARAT. Tel: (0260)2432716, 2401718 & 19 Fax: (0260)2431823

Corporate Office : UPL House, CTS No.610 B/2, Off. Western Express Highway, Bandra (East), Mumbai - 400051. Tel : 91-22-71528000

Marketing Office : Ready Money Terrace, 167, Dr. A B Road, Worli, Mumbai-400 018. India. Tel :91-22-2495 4628, 2493 0681, Fax : 91-22-24978119